



**CSIR-INDIAN INSTITUTE OF INTEGRATIVE MEDICINE**

(Formerly known as RRL)

Council of Scientific & Industrial Research

**CANAL ROAD, JAMMU**

Phones: 0191-2585000, 2585006 to 13, 15, 18 (11 Lines) Tele fax: 2585032, 2585019

File No. 06(159)/19-P

Dated: 28.08.2019

**CORRIGENDUM**

Reference to Tender\_id: 2019\_CSIR\_30538\_1 for **Upright freezer with stabilizer**, it is to notify that the some minor changes in specification of equipment is attached herewith as **Annexure-A**.

Kindly quote accordingly.

Sd/-

Stores & Purchase Officer

**Date: 28.08.2019**

**Revised specifications of Upright Deep Freezer with stabilizer with reference to Tender\_id:  
2019\_CSIR\_30538\_1**

- Upright freezer
- 650-750L,
- Temperature -86°C;
- Internally should be of Stainless steel;
- Insulation: vacuum/polyurethane Foam/their combination
- Freezer must have HFC-free, CFC-FREE, HCFC-FREE
- Warm up time 4-6 h
- Alarm: High/low temperature, power fail, filter clean fault
- lockable outer door
- System should have two compressors.
- System must have digital control panel
- IEC 61010 for Electrical safety (certificate should be attached)
- Freezer must be USFDA OR Certified CE from European Body OR BIS Certified
- Compartment            3-4
- Warranty: - 3 years on System & Servo stabilizer of 5 KV.
- System should have Silence level/Noise level 53 dB or lesser
- Freezer should have the minimum sample (2" vials) capacity of 50,000 or more
- Pull down time not >than 6h
- Warm up time >40h
- All the features mentioned above should be clearly highlighted in the Boucher provided as hard copy and should available in online broucher as well.
- Instrument should reach the institute within 45 days after getting the order.