

By hand

सी एस आई आर - भारतीय समवेत औषध संस्थान, जम्मू
(पूर्व में आर.आर.एल.)
(काउंसिल ऑफ साईटिफिक अँड इंडस्ट्रियल रिसर्च)
कनाल रोड जम्मू - 180 001
फोन: 0191-2569000, 2569001 फेक्स : 2569023, 2569019
Email: praphul@iiim.res.in/purchase@iiim.ac.in

दिनांक : 04.01.2021


No.06 (280)/19-पी
M/s Sigma Trading Company,
Shaheedi Chowk, Residency Road,
Jammu - 180 001.
श्रीमान/श्रीमती,

कृपया नीचे दिए गए नियमों एवं शर्तों के आधार पर निम्नलिखित सामग्री(यों) की आपूर्ति करने की व्यवस्था करें।
आपूर्ति की तारीख का सख्ती से पालन किया जाना चाहिए अन्यथा आपूर्ति आदेश रद्द किया जा सकता है।

S.No.	Description of the item	Qty.	Rate/unit	Amount INR
1.	Bench Top Refrigerated Centrifuge UD FDA APPROVED; Make : SORVALL; MODEL : ST 8R WITH ROTORS & STABILISER .	4	595,000/-	23,80,000.00 /
				.Total 23,80,000.00
SPECIFICATIONS :				
Type of Motor	With Carbon Brushless			
Configuration of Rotor supplied with centrifuge machine	Swing out and Angle rotor			
Condensed Water Drainage system	Yes			
Servo Controller Voltage stabilizer supplied with item	Yes			
Size of Rotor supplied with centrifuge in M1	24x1.5ml (for angle rotor) RPM 160000 or more 24x2 ml (FOR ANGLE ROTOR rpm 16000 2 Micro plate of 96 well each (for swing out rotor or angle rotor type rpm 3500) 8x15 ml (for angle rotor) rpm 13000 6 X 50ML ML (FOR ANGLE ROTOR) rpm 13000			
Warranty	2 years			
No. of acceleration and deceleration profiles in built for rotor speed while automatic breaking	Acceleration and deceleration time to be less than 15 seconds.			
Type of lid Lock	Electronic lid locking with automatic lid opening at end of run			
Maximum Timer set Point	60 minutes			
Conformity to standards	European CE (Notified Body) with 4 digit notified Body No.			
Accessories	Spanner to tight or loose Nut of Rotor on Motor shaft head puller, spare rotor nuts, fuse-2, instruction Manual Book, installation CD.			
Max. RPM with no load	20000			
Type of lid lock	Electronic lid locking with automatic lid opening at end of run			
Temperature Range	-20 to + 40 Celsius (Minus 10 to Plus 40)			
Imbalance Detection system	Continuous vibration measurement, with rotor mass correction			
Programs	Should have more than 50 programs			
Centrifugation Chamber	Stainless Steel			
Max. Timer Range	9h, 99min + continuous			
Certifications:	International/European CE, ISO13485-IVD			
Urgent Additional Terms and Condition	Micro plate Rotor for 4 or more plates with RCF of 2500 x g or more; Machine should be USFDA approved			

नियम और शर्तें /Terms and conditions

एफओआर आईआईआईएम केंद्रीय भंडार जम्मू /For IIIM Central Store, Jammu/Srinagar.
जीएसटी %5 अतिरिक्त/GST 5% extra. आपूर्ति 6-8 सप्ताह के भीतर./Delivery 6-8 weeks time.
Payment terms : Against delivery or ILC
Warranty : 2 years from the date of Installation.

आपका आभारी,

भंडार एवं क्रय अधिकारी

कॉपी:-

1. मांगकर्ता - Dr.G.D.Singh, Scientist.
2. वित्त एवं लेखा अनुभाग :निदेशक सक्षम/प्राधिकारी, आईआईआईएम, जम्मू ने, M/s Sigma Trading Company, Jammu की लागत के लिए रुपये 25,00,000/- की सहर्ष व्यय मंजूरी दे दी है। बजट कोड MLP-110014 से शामिल व्यय को विकलन किया जाना है।
3. केन्द्रीय स्टोर 4. मास्टर फाइल.